

Narrow implants

The possibility of placing implants can sometimes be limited due to physical conditions, e.g. where the horizontal space is limited by adjacent teeth and roots, or in situations with a narrow alveolar ridge. By using a narrow implant the need for bone augmentation or orthodontic tooth movement can be avoided. In situations with limited horizontal space a narrow diameter implant may be the only option to replace a missing tooth.

Several studies evaluating the clinical outcome of narrow implants (<3.5 mm in diameter) in general, placed in different indications, are available. Narrow implants supporting single tooth replacements have shown favorable clinical results¹⁻⁹ in the long-term perspective^{4,5,7-9}. Moreover, studies evaluating fixed partial dentures have shown good clinical results both after short-¹⁰ and long-term follow-up periods^{7-9,11}, even when using an immediate loading protocol¹². Narrow implants have also been used to support full arch reconstructions, and satisfactory results have been shown for fixed bridges⁹ and overdentures in the mandible^{9,13,14} and in the maxilla^{9,15}. In general, no difference in the clinical outcome between standard diameter implants and narrow implants has been observed^{1,3,7,16-19}. In an extensive review, Renouard and Nisand concluded that survival rates for narrow implants are comparable with that of standard diameter implants when used in appropriate indications. They also reported that no relationship was found between marginal bone loss and implant diameter²⁰.

Results from a study specifically evaluating Astra Tech implants with a diameter of 3.5 mm, placed in the atrophic maxilla, shows high survival rate and a mean marginal bone level reduction of 0.3 mm after one year²¹.

The most narrow implant developed by Astra Tech, OsseoSpeed™ 3.0 S, has a diameter of 3.0 mm and the same features as the regular OsseoSpeed™ implant. OsseoSpeed™ 3.0 S is indicated for different surgical techniques and loading protocols when replacing maxillary lateral incisors and mandibular lateral and central incisors.

Published data indicates that treatment with OsseoSpeed™ 3.0 S implants is safe and predictable in sites with limited physical space in anterior regions²²⁻²⁴.

1. Andersen E, Saxegaard E, Knutsen BM, Haanaes HR. A prospective clinical study evaluating the safety and effectiveness of narrow-diameter threaded implants in the anterior region of the maxilla. *Int J Oral Maxillofac Implants* 2001;16(2):217-24. [Abstract in PubMed](#)
2. Cordaro L, Torsello F, Mirisola Di Torresanto V, Rossini C. Retrospective evaluation of mandibular incisor replacement with narrow neck implants. *Clin Oral Implants Res* 2006;17(6):730-5. [Abstract in PubMed](#)
3. Mericske-Stern R, Grutter L, Rosch R, Mericske E. Clinical evaluation and prosthetic complications of single tooth replacements by non-submerged implants. *Clin Oral Implants Res* 2001;12(4):309-18. [Abstract in PubMed](#)
4. Polizzi G, Fabbro S, Furri M, Herrmann I, Squarzonì S. Clinical application of narrow Branemark System implants for single-tooth restorations. *Int J Oral Maxillofac Implants* 1999;14(4):496-503. [Abstract in PubMed](#)
5. Vigolo P, Givani A. Clinical evaluation of single-tooth mini-implant restorations: a five-year retrospective study. *J Prosthet Dent* 2000;84(1):50-4. [Abstract in PubMed](#)
6. Zarone F, Sorrentino R, Vaccaro F, Russo S. Prosthetic treatment of maxillary lateral incisor agenesis with osseointegrated implants: a 24-39-month prospective clinical study. *Clin Oral Implants Res* 2006;17(1):94-101. [Abstract in PubMed](#)
7. Romeo E, Lops D, Amorfini L, Chiapasco M, Ghisolfi M, Vogel G. Clinical and radiographic evaluation of small-diameter (3.3-mm) implants followed for 1-7 years: a longitudinal study. *Clin Oral Implants Res* 2006;17(2):139-48. [Abstract in PubMed](#)
8. Vigolo P, Givani A, Majzoub Z, Cordioli G. Clinical evaluation of small-diameter implants in single-tooth and multiple-implant restorations: a 7-year retrospective study. *Int J Oral Maxillofac Implants* 2004;19(5):703-9. [Abstract in PubMed](#)
9. Zinsli B, Sagesser T, Mericske E, Mericske-Stern R. Clinical evaluation of small-diameter ITI implants: a prospective study. *Int J Oral Maxillofac Implants* 2004;19(1):92-9. [Abstract in PubMed](#)
10. Hallman M. A prospective study of treatment of severely resorbed maxillae with narrow nonsubmerged implants: results after 1 year of loading. *Int J Oral Maxillofac Implants* 2001;16(5):731-6. [Abstract in PubMed](#)
11. Comfort MB, Chu FC, Chai J, Wat PY, Chow TW. A 5-year prospective study on small diameter screw-shaped oral implants. *J Oral Rehabil* 2005;32(5):341-5. [Abstract in PubMed](#)
12. Degidi M, Nardi D, Piattelli A. Immediate restoration of small-diameter implants in cases of partial posterior edentulism: a 4-year case series. *J Periodontol* 2009;80(6):1006-12. [Abstract in PubMed](#)
13. Cho SC, Froum S, Tai CH, Cho YS, Elian N, Tarnow DP. Immediate loading of narrow-diameter implants with overdentures in severely atrophic mandibles. *Pract Proced Aesthet Dent* 2007;19(3):167-74. [Abstract in PubMed](#)
14. Spiekermann H, Jansen VK, Richter EJ. A 10-year follow-up study of IMZ and TPS implants in the edentulous mandible using bar-retained overdentures. *Int J Oral Maxillofac Implants* 1995;10(2):231-43. [Abstract in PubMed](#)
15. Payne AG, Tawse-Smith A, Thomson WM, Duncan WD, Kumara R. One-stage surgery and early loading of three implants for maxillary overdentures: a 1-year report. *Clin Implant Dent Relat Res* 2004;6(2):61-74. [Abstract in PubMed](#)
16. Weng D, Jacobson Z, Tarnow D, Hurzeler MB, Faehn O, Sanavi F, et al. A prospective multicenter clinical trial of 3i machined-surface implants: results after 6 years of follow-up. *Int J Oral Maxillofac Implants* 2003;18(3):417-23. [Abstract in PubMed](#)
17. Davarpanah M, Martinez H, Tecucianu JF, Celletti R, Lazzara R. Small-diameter implants: indications and contraindications. *J Esthet Dent* 2000;12(4):186-94. [Abstract in PubMed](#)
18. Block MS, Kent JN. Cylindrical HA-coated implants-8-year observations. *Compend Suppl* 1993(15):S526-32; quiz S65-6. [Abstract in PubMed](#)
19. Degidi M, Piattelli A, Carinci F. Clinical outcome of narrow diameter implants: a retrospective study of 510 implants. *J Periodontol* 2008;79(1):49-54. [Abstract in PubMed](#)
20. Renouard F, Nisand D. Impact of implant length and diameter on survival rates. *Clin Oral Implants Res* 2006;17 Suppl 2:35-51. [Abstract in PubMed](#)
21. Veltri M, Ferrari M, Balleri P. One-year outcome of narrow diameter blasted implants for rehabilitation of maxillas with knife-edge resorption. *Clin Oral Implants Res* 2008;19(10):1069-73. [Abstract in PubMed](#)
22. Galindo-Moreno P, Nilsson P, King P, Worsaae N, Maiorana C, Schramm A. Evaluation of OsseoSpeed™ 3.0 S implants, a 6 months follow-up. Paper presented at: Academy of Osseointegration, 2010; March 4-6, Orlando, Florida.
23. Martin R. Astra Tech OsseoSpeed™ 3.0S implant. *Inside Dentistry* 2010;6(4):2-4.
24. Caudry S, Landzberg M. Clinical considerations for the use of a two-piece 3.0 mm implant for replacing a missing mandibular incisor. *Journal of Clinical & Practical Oral Implantology* 2010;1(2):25-29.

